

# Product Safety Recall



## Rite Aid Mini Digital Temple Touch Thermometer LOT NUMBER: 5216

Sold at Big W, TerryWhite Chemmart, Toys R Us,  
and selected local pharmacies nationally  
(June 2017 - February 2018)

**Defect** The battery compartment on the thermometer (lot number: 5216) is not adequately secured, allowing access to the button battery inside.

**Hazard** If young children gain access to the button batteries and ingest them, they may suffer internal burn injuries, which can result in serious illness and even death. In addition, the batteries may pose a choking hazard to young children.

**What to do** Consumers should immediately stop using the thermometer and return it to below reply-paid address for an exchange.

ToLife Technologies  
Reply Paid 36  
PO Box 36  
Welshpool, WA 6986

**Contact Details** You can contact ToLife Technologies by phoning +61 8 9350 5688, Mon-Fri 8:00am-5:00pm AWST, via e-mail at [marketing@tolife.com.au](mailto:marketing@tolife.com.au), or by visiting [www.riteaid.com.au](http://www.riteaid.com.au)

**See [productsafety.gov.au](http://productsafety.gov.au) for  
Australian product recall information**